

UH/SLS/Biochem/KM/RS-1

Sealed quotations are invited by Department of Biochemistry, School of Life Sciences, University of Hyderabad from reputed manufacturers or their authorized agents for the **supply and installation of**

**Biosafety cabinet (class II), A2 compliant for all features. This includes**

- Class II (A2) with exhaust through HEPA filters only
- 70% recirculation (through HEPA filters) with 30% exhaust
- Work Size of the hood 3 ftWx2 ftDx2 ft H
- UV lamp
- HEPA filters with 0.3 micron and 99.9% efficiency.
- Pre filters for microparticulates
- The unit should provide vertical air flow
- Air Inflow Rate should be at least 100 fpm
- Exhaust duct to send out air
- Front sash with sliding mechanism
- LED light illumination
- Stainless steel work surface
- Pressure gauge with display
- Rust free construction
- Liquid drain off collection

**Terms and Conditions:**

- 1) Warranty period should be clearly mentioned. After sales terms of services should be clearly given. Enclosure of 'user-list' is preferred
- 2) Delivery should be upto School of Life Sciences, University of Hyderabad. Basic cost, taxes, duties, packing & forwarding charges, transport & insurance charges etc. should be separately indicated wherever applicable.
- 3) Validity: The rates quoted should be firm till the date of satisfactory execution of the order.
- 4) Free of cost installation
- 5) Mode of Payment: 100% payment against satisfactory installation.
- 6) Firms should mention their VAT, Service Tax, PAN registration numbers etc. on the quotations.
- 7) No part quotations shall be accepted/processed.all material required for installation must be clearly quoted for.
- 8) The acceptance of the quotation(s) will rest solely with the University of Hyderabad who is not bound to accept the lowest quotation and reserves the right to reject oraccept any or all the quotations received without assigning any reason. Its decision in the matter will be final and binding on all the tenders.
- 9) Other terms & conditions will be as per University of Hyderabad purchase procedure in vogue and as amended from time to time.
- 10) Important: Please attach a summary table to highlight the compatibility of your equipment's set-up against specific features mentioned here, with adequate documentation support. Quotations will not be considered if the table is missing.
- 11) Sealed envelopes, clearly marked "**Quotation for Biosafety cabinet**" may be sent to the following address:

Dr. Renu Shukla,  
C/o Prof. Krishnaveni Mishra  
# F-59, Department of Biochemistry,  
School of Life Sciences, South Campus  
University of Hyderabad,  
Gachibowli, Hyderabad 500046,  
Telangana, India

Quotations should reach the above address by 5.00PM on 15<sup>th</sup> February 2020.